




User manual

SKB037(B) **Stainless steel patient trolley**



SAIKANG / Medical

Jiangsu Saikang Medical Equipment Co., Ltd.

STOCK CODE: 870098

Contents

Manual introduction	2
1. Installing guide	3
1.1 List of standard accessories	3
1.2 Install casters	5
1.3 Install the feet.....	6
1.4 Install the bed board.....	7
2. Product instructions.....	9
2.1 Use method of casters	10
2.2 Use method of guardrails	11
2.3 Backboard instructions	12
2.4 IV pole installation	13
2.4.2 Usage of the IV Pole.....	14
3. Faults exclusion	14
4. Stainless steel patient trolley maintenance.....	14
5. Packaging, transport, storage.....	15
6. After-sales service	16

Manual introduction

Note:

Please carefully read this manual and strictly follow the steps before installing and using. We require all users to have the knowledge of user's manual and all the cautions.

► This manual is applicable to SKB037(B) Stainless steel patient trolley. It contains the installation method, operating instructions and maintenance check points.

► The graphics in this manual are for users' reference only. In actual use, please refer to the real object.

► This user manual includes the description, instruction, maintenance and technical parameters of the product.

► Any questions, please contact the after-sale service center or distributor.

► Anyone using or operating this product should read and comply with it.

► Except for some unexpected situations, users are required to be familiar with the usage and precautions.

1. Installing guide

1.1 List of standard accessories

After opening the packing case, please check whether the following parts are complete in the packing case:

Table a:














NO.	Name	Picture	Specification	QTY	Unit
1	Foot right		L1600*W485mm	1	pc
2	Foot left		L1600*W485mm	1	pc
3	Foot connection tube		Φ32*551-M8mm	2	pcs
4	The foot connecting tube has holes		Φ32*551-M8mm	2	pcs
5	Bed frame		L1960*W520mm	1	pc
6	Casters		5-inch cover wheel screw with brake	4	pcs
7	IV pole		Φ16-19mm/ off white	1	pc
8	Screw	/	Refer to Table a.1	1	set


Table a.1:

NO.	Name	Picture	Specification	QTY	Unit
1	Spring washer		$\Phi 16 * \Phi 23.8 * H3.5mm$	4	pcs
2	Hexagon Screw		M8*70mm	4	pcs
3	Flat washer		$\Phi 8 * \Phi 16 * H1.2mm$	4	pcs
4	Spring washer		$\Phi 8 * \Phi 14 * H2.5mm$	4	pcs
5	Lock nut		M8mm	4	pcs
6	Hexagon socket screw		M8*45mm	8	pcs
7	Flat washer		$\Phi 8 * \Phi 16 * H1.2mm$	8	pcs
8	Spring washer		$\Phi 8 * \Phi 14 * H2.5mm$	8	pcs

1.2 Install casters



Lay a protective pad and put the bed body upside down;

Take out the $\Phi 16$ spring washer from the screw pack, take out the cover wheel, insert the cover wheel into the tube and tighten it clockwise, and then tighten it with an hex wrench.





1.3 Install the feet

Fix the foot with an Allen wrench as shown in the figure.





1.4 Install the bed board

Put the bed upright, lift the backboard, take out the screws, and install as shown.




Lay a layer of protective cushions and turn the main body of the bed upside down; remove the screws and install as shown.

(In order to facilitate observation, temporarily hide the bed board.)




2. Product instructions




Dimensions	Length	mm	1960
	Width		520
	Height from the ground		770
Weight	Safe working load	KG	170

2.1 Use method of casters

Lock the caster: After pushing the bed to the desired position, press the ON button with your foot and lock the caster when you hear a “click”, as shown in the figure below.




Unlock the caster: Press the OFF button with your foot and unlock the castor when you hear a “click”.




2.2 Use method of guardrails

Raise the guardrail, when you hear the sound of “click”, it means the guardrail is locked.




Lower the guardrail, pull the guardrail back in the spring direction as shown in the figure, and then lower the guardrail.





2.3 Backboard instructions

Lift the backboard and release it to the designated position to automatically fix the position.



Raise the backboard to its maximum height and then lower it to lower the back.






2.4 IV pole installation

Female part installation: Press the plastic hook (female part) down on the $\Phi 16\text{mm}$ stainless steel pipe.

Male part installation: Place the plastic hook (male part) directly above the female part in a cross shapes, and then aligns the buckles on the male part with the slots on the female part and press down. After hearing a "click", check whether the end of the buckle is tight, if it is tight, then complete the installation, if not, continue to press down until the buckle is tight.

Insert the installed IV pole vertically into the IV pole prevision. (The arrow in the picture shows the IV pole prevision)




2.4. 2 Use method of IV pole

Adjusting the height of the IV pole:

1. Loosen the outer sleeve counter clock carefully, insert the inner tube of the IV pole into the outer sleeve, and then adjust the length of the extension rod up and down;
2. Adjust the required length, and then tighten the outer sleeve clock carefully. The maximum bearing of each hook is 2Kg, a total of 8Kg.

(Height Adjustment :890--1540mm)



3. Faults exclusion

Common failures	Probable reasons	Exclusion methods
Difficult to move the whole patient trolley	1. The caster is locked without releasing the lever 2. The caster is damaged	1. Release the brake,check 2.1 2.Replace the caster,check1.2

4. Stainless steel patient trolley maintenance

1.In order to use the patient trolley safely, it is necessary to perform regular safety inspections on the patient trolley. It is recommended to check it every six months to ensure that the connections

are not loose and all operations are normal.

2. When the patient trolley frame ages and it reaches a certain servicing time, which the metal part of the patient trolley and plastic parts of the patient trolley frame can be recycle.

3. Avoid scratching the panel with sharp-angle appliances or knives during use, and clean it frequently to keep it clean and dry.

4. If the panel is accidentally stained with stains, it is recommended to clean it. Do not use alkaline or corrosive chemicals to clean the patient trolley, which will cause the stainless steel surface to rust.

5. Regularly check the casters to avoid hitting or overloading, which will cause damage to the casters.

6. It is forbidden to force the casters to brake when braking, which will cause damage to the brake system.

7. The wear of the tire surface of the wheel can be detected by visual inspection. Some thread spun yarn and other accumulations may be wrapped around the wheel, remove the bolts and nuts of the wheel, clean up the debris, and check whether the bearing of the wheel is damaged. If the parts are not damaged, can be reassembled and used.

8. Judging whether the caster is replaced according to the time used, wear and tear, you can replace the wheels and make sure that the axles bolt and nuts are well put together and connected. If possible, use a lock washers or locknuts to help you with this process.

9. For the casters equipped with brakes, it is necessary to regularly check whether the brakes are working properly. Check the brakes once a day or before each use. For a patient trolley equipped with multiple brake casters, only one brake caster can be locked at a time which will allow you to try and push the patient trolley also check whether one of the brake performance of each caster is good. If the function of the brake fails due to wear or damage to the wheels, immediately replace the wheels and then retest the brake system.

10. If the mechanism of the brake system of the caster is damaged, and the brake needs to be repaired or replaced, please contact our after-sales service department or authorized dealer provide by our company. Every time the brakes are replaced, the brake performance of the casters must be retested.

5. Packaging, transport, storage

1. The packaging of stainless steel patient trolley is carried out according to the contract or product standard.

2. Avoid rushing out, violent vibration, and protection from sunlight and rain during the transportation of the stainless steel patient trolley.

3. The stainless steel patient trolley should be stored:

a) Ambient temperature: $-40^{\circ}\text{C} \sim 70^{\circ}\text{C}$.

b) Relative humidity: $\leq 95\%$.

c) Atmospheric pressure: $500\text{hPa} \sim 1060\text{ hPa}$.

6. After-sales service

1. Please keep the files that comes with the machine and invoices of this product properly, and you need to present these files when the company performs warranty and maintenance for the product.

2. If there is any problem in the process of using, please contact our company in time, so that our company can provide you with accurate and fast technical support and maintenance services in a timely manner.

3. From the date of sale, if the product is broken or damaged due to the correct installation and use according to the specification, the product will enjoy one-year free warranty and lifelong maintenance service with the "certificate" or invoice.

4. From the date of purchase, if it is indeed damaged or does not work normally due to quality problems within one year, the company will provide to the users a free repairing of the product.

5. Lifetime service from manufacturer: JIANGSU SAIKANG MEDICAL EQUIPMENT CO LTD



MODEL:SKB037(B) | REF.: 319-SKB037(B)

CAMILLA / STRETCHER / Civière / BARELLA / MACA / TRAGEBRETT



Jiangsu Saikang Medical Equipment Co.,Ltd

No. 35 Lehong Road, Modern Agriculture
Demonstration Park, Zhangjiagang City,
Jiangsu Province, China



SUNGO Europe B.V.

Fascinatio Boulevard 522, Unit 1.7,
2909VA Capelle aan den IJssel,
The Netherlands



Quirumed S.L.U. CIF B97267405

C/Corretger 117 Pq.Emp.Tactica,Paterna,
46980, Valencia(Spain)

LOT:HW25xxx



2025-xx-xx



Jiangsu Saikang Medical Equipment Co., Ltd.

STOCK CODE: 870098



No. 35, Lehong Road · Zhangjiagang (Jiangsu) · China

+86 18021231901

sales@chinasaikang.com

www.saikangmedical.com